



**DIRECTORATE OF RESEARCH SERVICES**  
**INDIRA GANDHI KRISHI VISHWAVIDYALAYA**  
KRISHAK NAGAR, RAIPUR 492 012 (C.G.)  
Phone & Fax: 0771-2443035


No.DRS/Store/2017/SSAE/3280


Raipur, Dated: 15/02/2017

**CORRIGENDUM**

With reference to tender release vide No.DRS/Store/2017/3240 dated 10/02/2017 for supply of "*Inductively Coupled Plasma Emission spectrometer*" the condition given in point No. f, h. and k. has been ratified , these specifications may please be read as :-

- f. **Optics** -High Speed dual Monochromator/ polychromator based spectrometer with Echelle Grating and optical resolution of minimum 0.008 nm or better at 200 nm. Wavelength Range -170-900 nm.
- h. **Detector**- UV- Sensitive, dual back side-illuminated Charge- Coupled Device CCD/CID/CMOS array detector having cooling facility using a single-stage intergrated cooling operated at approximately- 8<sup>0</sup>C. Stabilize the instrument in less than 15 minnutes.
- k. **Accessories**- Organic Sample, HF Sample analysis kit, Hydride generation accessory kit, Chiller from manufacturer shall be supplied along with basic system, Standard solution for calibration/validation to be mentioned separately in the offer and others.

  
15/2/2017  
Director of Research

  
15/2/17  
Professor & Head  
Department of Soil Science  
Indira Gandhi Krishi Vishwavidyalaya  
Raipur - 492009 (C.G.)